

Bid Document

| Bid Details | |
|--|---|
| Bid End Date/Time | 08-07-2022 17:00:00 |
| Bid Opening Date/Time | 08-07-2022 17:30:00 |
| Bid Offer Validity (From End Date) | 80 (Days) |
| Ministry/State Name | Ministry Of Science And Technology |
| Department Name | Department Of Science And Technology (dst) |
| Organisation Name | Sree Chitra Tirunal Institute For Medical Sciences And Technology (sctimst) |
| Office Name | Thiruvananthapuram |
| Total Quantity | 100 |
| Item Category | Smooth bore Circuits- Anesthesia Breathing Circuits (Q3) |
| MSE Exemption for Years Of Experience and Turnover | Yes |
| Startup Exemption for Years Of Experience and Turnover | Yes |
| Document required from seller | OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| Bid to RA enabled | No |
| Time allowed for Technical Clarifications during technical evaluation | 2 Days |
| Evaluation Method | Total value wise evaluation |

EMD Detail

| | |
|----------|----|
| Required | No |
|----------|----|

ePBG Detail

| | |
|----------|----|
| Required | No |
|----------|----|

Splitting

Bid splitting not applied.

MII Purchase Preference

| | |
|-------------------------|----|
| MII Purchase Preference | No |
|-------------------------|----|

MSE Purchase Preference

| | |
|-------------------------|-----|
| MSE Purchase Preference | Yes |
|-------------------------|-----|

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria". If the bidder is OEM of the offered products, it would also be exempted from the "OEM Average Turnover" criteria. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria". If the bidder is OEM of the offered products, it would also be exempted from the "OEM Average Turnover" criteria. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.

Smooth Bore Circuits- Anesthesia Breathing Circuits (100 pieces)

| | |
|------------|------------------|
| Brand Type | Registered Brand |
|------------|------------------|

Technical Specifications

* As per GeM Category Specification

| Specification | Specification Name | Bid Requirement (Allowed Values) |
|---------------|--|----------------------------------|
| GENERAL | Usage | Adults |
| | Utility | Single use- disposable |
| DIMENSIONS | Corrugated Tube ID(mm) | 22 |
| | Corrugated Tube length (mm) | 1600 |
| | Fresh Gas flow tube dia (mm) | 22 |
| | Fresh Gas flow tube length (mm) | 1600 |
| | Reservoir bag size to allow respiratory monitoring and /or assistance (ml) | 1000 |
| MATERIAL | Material of the reservoir bag | Latex free |

| Specification | Specification Name | Bid Requirement (Allowed Values) |
|---------------|---|---|
| CONSTRUCTION | Reservoir bag | Not available |
| | Reservoir bag type | Open Tail |
| | Fixed elbow and luer lock port for safe monitoring at the patient end | Yes |
| | Expiratory valve at the patient end | Yes |
| | Double swivel elbow elastomeric cap at patient end | Yes |
| | Accurate and adjustable pressure limiting valve with safety feature | Yes |
| | Pressure adjustment range of pressure limiting value (kg/cm2) | 3 |
| | Suitable adaptors to connect to various parts of breathing system, flow meter etc | Yes |
| | Y-piece connector with port | Yes |
| | Humidification limb | Yes |
| | Facility for incorporating inline Nebulizer/MDI | Yes |
| PACKING | Type of packing | Single, individually packed in blister pack |

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days |
|-------|-----------------------------|--|----------|---------------|
| 1 | Shiju V S | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum | 100 | 15 |

Buyer Added Bid Specific Terms and Conditions

1. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

2. **Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

3. **Certificates**

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

4. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

5. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

6. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

7. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

8. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity/restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and/or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents/clauses shall also be null and void. If any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations. Also, GeM does not permit collection of Tender fee / Auction fee in case of Bids / Forward Auction as the case may be. Any stipulation by the Buyer seeking payment of Tender Fee / Auction fee through ATC clauses would be treated as null and void.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---