

Bid Document

Bid Details	
Bid End Date/Time	28-07-2022 17:00:00
Bid Opening Date/Time	28-07-2022 17:30:00
Bid Offer Validity (From End Date)	80 (Days)
Ministry/State Name	Ministry Of Science And Technology
Department Name	Department Of Science And Technology (dst)
Organisation Name	Sree Chitra Tirunal Institute For Medical Sciences And Technology (sctimst)
Office Name	Thiruvananthapuram
Total Quantity	50
Item Category	Smooth bore Circuits- Anesthesia Breathing Circuits (Q3)
MSE Exemption for Years Of Experience and Turnover	Yes
Startup Exemption for Years Of Experience and Turnover	Yes
Document required from seller	OEM Authorization Certificate,Additional Doc 1 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Time allowed for Technical Clarifications during technical evaluation	2 Days
Evaluation Method	Total value wise evaluation

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

MII Purchase Preference

MII Purchase Preference	No
-------------------------	----

MSE Purchase Preference

MSE Purchase Preference	Yes
-------------------------	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria". If the bidder is OEM of the offered products, it would also be exempted from the "OEM Average Turnover" criteria. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria". If the bidder is OEM of the offered products, it would also be exempted from the "OEM Average Turnover" criteria. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.

Smooth Bore Circuits- Anesthesia Breathing Circuits (50 pieces)

Brand Type	Registered Brand
------------	------------------

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Bid Requirement (Allowed Values)
GENERAL	Usage	Adults
	Utility	Single use- disposable
DIMENSIONS	Corrugated Tube ID(mm)	22
	Corrugated Tube length (mm)	1600
	Fresh Gas flow tube dia (mm)	22
	Fresh Gas flow tube length (mm)	1600
	Reservoir bag size to allow respiratory monitoring and /or assistance (ml)	1000

Specification	Specification Name	Bid Requirement (Allowed Values)
MATERIAL	Material of the reservoir bag	Latex free
CONSTRUCTION	Reservoir bag	Not available
	Reservoir bag type	Open Tail
	Fixed elbow and luer lock port for safe monitoring at the patient end	Yes
	Expiratory valve at the patient end	Yes
	Double swivel elbow elastomeric cap at patient end	Yes
	Accurate and adjustable pressure limiting valve with safety feature	Yes
	Pressure adjustment range of pressure limiting value (kg/cm2)	3
	Suitable adaptors to connect to various parts of breathing system, flow meter etc	Yes
	Y-piece connector with port	Yes
	Humidification limb	Yes
	Facility for incorporating inline Nebulizer/MDI	Yes
PACKING	Type of packing	Single, individually packed in blister pack

Additional Specification Parameters - Smooth Bore Circuits- Anesthesia Breathing Circuits (50 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
1	ADULT/PED BREATHING CIRCUIT, 6 FT PATIENT CIRCUIT INTERNAL DIAMETER: 22MM, RINSP@ 30LPM : 0.02HPA/L/MIN REXP@ 30LPM: 0.10 HPA/L/MIN 4. CT@ 60 HPA: 2.8 ML/HPA 5. DEAD SPACE: 22 ML MAXIMUM WORKING PRESSURE: 100HPA (CM H2O) COMPLIES WITH ISO 5367.
2	DISPOSABLE ADULT SINGLE LIMB , CORRUGATED HOSE , WYE PATIENT CIRCUIT. CORRUGATED HOSE (LONG) LENGTH: 6FT CORRUGATED HOSE (SHORT) LENGTH: 1.1FT INTERNAL DIAMETER (ID) : 22MM

Specification Parameter Name	Bid Requirement (Allowed Values)
3	1.83M VENTILATOR WYE-- 1.5MM INTERNAL DIAMETER (TO PATIENT SIDE) AND 2.3MM INTERNAL DIAMETER (TO TUBING SIDE) , CORRUGATED BREATHING CIRCUIT WITH CUT TO SIZE TUBE AND EXHALATION VALVE(HAVING SILICON DIAPHRAGM).
4.	EXHALATION DRIVE LINE THICK WALL TUBE (LENGTH : 6FT) WITH WIDE CONNECTOR.(1.8M) PRESSURE LINE(GREEN)THICK WALL TUBE(LENGTH : 6FT) WITH WIDE CONNECTOR.(1.8M
5	BIFURCATED WYE WITH PATIENT CONNETOR: (22MM OUTER DIAMETER AND 15MM INNER DIAMETER) AND AIR WAY PRESSURE LINE PORT.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	50	15

Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

3. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

4. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

1.Provide Product catalog.

2.Sample must be submitted for technical evaluation Sample must be carefully packed sealed and labeled ,with GeM bid number, Item Name and senders name for easy identification. Sample should be reached at "Sr.Purchase &Stores Officer,AMCHSS,SCTIMST,Kerala,TVM-11 within

in 3 days of Bid End date /Bid Opening date, otherwise offer will be rejected.

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity/restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and/or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents/clauses shall also be null and void. If any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations. Also, GeM does not permit collection of Tender fee / Auction fee in case of Bids / Forward Auction as the case may be. Any stipulation by the Buyer seeking payment of Tender Fee / Auction fee through ATC clauses would be treated as null and void.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---