

CORRIGENDRUM/ADDENDUM
ELECTRO SURGICAL UNIT WITH BIPOLAR (TENDER NO: SCT/004/IMP/20171-18/SA5)

EXISTING	TO BE READ AS
<ul style="list-style-type: none">• SHOULD HAVE TWO BIPOLAR SOCKETS AND ONE MONOPOLAR SOCKET, SHOULD HAVE INTERNAL COOLING SYSTEM	<ul style="list-style-type: none">• SHOULD HAVE ONE BIPOLAR SOCKET AND TWO MONOPOLAR SOCKET, SHOULD HAVE INTERNAL COOLING SYSTEM. TWO BIPOLAR SOCKETS WILL BE A PREFERRABLE FEATURE.
<ul style="list-style-type: none">• MONO POLAR FORCEPS-2 NOS	<ul style="list-style-type: none">• MONOPOLAR FORCEPS/PENCIL-2 NOS
<ul style="list-style-type: none">• Last date of receipt of tender: 17/10/2017 upto 01.00PM	<ul style="list-style-type: none">• Last date of receipt of tender 02/11/2017 upto 01.00PM
<ul style="list-style-type: none">• Tender will be opened 19/10/2017 at 03.00PM	<ul style="list-style-type: none">• Tender will be opened at 02/11/2017 at 03.00PM